

Framework No: 4526698
Framework Title: Medical and Surgical Sundries
Framework Period: 2 years from date of award
 (With the option to extend for any period(s) up to and including 24 months)

Product Specifications

This Tender is for the following Lots:

<u>Lot 1 – Medicine Measure 60ml</u>		Benchtop Evaluation Pass/Fail
1	Medicine Measure	
2	60ml	
3	Polypropylene	
4	Clear / Translucent – Measurement marking must be clear to read	
5	Graduated	
6	Single Use	
7	Non Sterile	
8	Will be subject to pressure test to ensure no cracking occurs	
9	Packaging must include; Product Code, Lot Number, Conformity Assessment Marking	

<u>Lot 2 – Jug 1 Litre</u>		Benchtop Evaluation Pass/Fail
1	Jug 1 Litre	
2	Graduated	
3	Polypropylene	
4	Opaque	
5	Single Wrapped	
6	Sterile	
7	Single Use	
8	Handle on jug must be sturdy to hold weight of fluid in jug	
9	Will be subject to test to ensure no cracking occurs	
10	Packaging must include; Sterile, Single Use, Product Code, Lot Number, Expiry Date, Conformity Assessment Marking	

<u>Lot 3 – Medicine Measure – 30ml</u>		Benchtop Evaluation Pass/Fail
1	Medicine Measure	
2	30ml	
3	Polypropylene	
4	Clear / Translucent – Measurement marking must be clear to read	

5	Graduated	
6	Single Use	
7	Non-Sterile	
8	Will be subject to test to ensure no cracking occurs	
9	Packaging must include; Single Use, Product Code, Lot Number, Conformity Assessment Marking	

<u>Lot 4 – Gallipot 120ml</u>		Benchtop Evaluation Pass/Fail
1	Gallipot 120ml	
2	Polypropylene	
3	Opaque	
4	Graduated	
5	Single Use	
6	Sterile	
7	Will be subject to pressure test to ensure no cracking occurs	
8	Packaging must include: Sterile, Single Use, Product code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 5 - Gallipot 250ml</u>		Benchtop Evaluation Pass/Fail
1	Gallipot 250ml	
2	Polypropylene	
3	Opaque	
4	Graduated	
5	Single Use	
6	Sterile	
7	Will be subject to pressure test to ensure no cracking occurs	
8	Packaging must include: Sterile, Single Use, Product code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 6 - Gallipot 60ml</u>		Benchtop Evaluation Pass/Fail
1	Gallipot 60ml	
2	Polypropylene	
3	Opaque	
4	Graduated	
5	Single Use	
6	Sterile	
7	Will be subject to pressure test to ensure no cracking occurs	
8	Packaging must include: Sterile, Single Use, Product code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 7 - Gallipot (Minimum 200ml – Maximum 300ml)</u>		Benchtop Evaluation Pass/Fail
1	Gallipot (Minimum 200ml – Maximum 300ml)	
2	Aluminium Foil	
3	Sterile	
4	Single Use	
5	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 8 – Scissors Suture Hook</u>		Benchtop Evaluation Pass/Fail
1	Scissors	
2	Suture Hook	
3	Stainless Steel	
4	11-13cm (Measurements full length from Handle to Tip)	
5	Single Use	
6	Sterile	
7	Will be tested to fit 8.5 glove wearer	
8	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 9 – Scissors Dressing Sharp</u>		Benchtop Evaluation Pass/Fail
1	Scissors Dressing	
2	Sharp	
3	Nickle Plated	
4	12-14cm (Measurements full length from Handle to Tip)	
5	Single Use	
6	Sterile	
7	Will be tested to fit 8.5 glove wearer	
8	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 10 – Pen Torch</u>		Benchtop Evaluation Pass/Fail
1	Pen Torch for medical use	
2	Reusable	
3	Non-Sterile	
4	Packaging must include: Product Code, Lot No, Conformity Assessment Marking	

<u>Lot 11 – Eye Shield</u>		Benchtop Evaluation Pass/Fail
1	Eye Shield	
2	Disposable	
3	Single Use	
4	Sterile	
5	Ventilation holes	
6	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 12 -Surgical Scrub Brush Dry Sponge</u>		Benchtop Evaluation Pass/Fail
1	Surgical Scrub Brush	
2	Dry Sponge	
3	Plastic Nail Cleaner	
4	Sterile	
5	Single Use	
6	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 13 – Surgical Scrub Brush with Chlorhexidine</u>		Benchtop Evaluation Pass/Fail
1	Surgical Scrub Brush	
2	Chlorhexidine	
3	Plastic Nail Cleaner	
4	Sterile	
5	Single Use	
6	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 14 – Surgical Scrub Brush with Iodophor</u>		Benchtop Evaluation Pass/Fail
1	Surgical Scrub Brush	
2	Iodophor	
3	Plastic Nail Cleaner	
4	Sterile	
5	Single Use	
6	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 15 – Dual Headed Stethoscope</u>		Benchtop Evaluation Pass/Fail
1	Dual Headed Stethoscope	
2	Approximately 80cm length	
3	Packaging must include Product Code, Lot No and Conformity Assessment Marking	

<u>Lot 16 – Dressing Forceps</u>		Benchtop Evaluation Pass/Fail
1	Dressing Forceps	
2	10-15cm approx. (Measurement is the full length)	
3	Single Use	
4	Sterile	
5	Will be grasp tested i.e. used to effectively and efficiently grasp swabs, cotton wool	
6	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 17 – Emergency Foil Blanket</u>		Benchtop Evaluation Pass/Fail
1	Emergency Foil Blanket	
2	Adult	
3	130cm x 210cm approx.	
4	Non-Sterile	
5	Individually Wrapped	
6	Packaging must include: Product Code, Lot No and Conformity Assessment Marking	

<u>Lot 18 – Tongue Depressor Wooden</u>		Benchtop Evaluation Pass/Fail
1	Tongue Depressor	
2	Wooden	
3	Non-Sterile	
4	Edges must be smooth to ensure product does not splinter	
5	Packaging must include: Product Code, Lot No and Conformity Assessment Marking	

<u>Lot 19 – Surgical Skin Markers Fine Tip</u>		Benchtop Evaluation Pass/Fail
1	Surgical Skin Marker	
2	Fine Tip	
3	Violet	
4	Low Smell	

5	Non-toxic	
6	Non-tattooing ink for all skin types	
7	Individually packed	
8	Single Use	
9	Sterile to come with separate 15cm ruler	
10	Suppliers or double ended marker – you will need to apply for both Fine Tip and Regular Lots if you wish to be considered and evaluated for both Lots	
11	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 20 – Surgical Skin Markers Regular Tip</u>		Benchtop Evaluation Pass/Fail
1	Surgical Skin Makers	
2	Regular Tip	
3	Violet	
4	Low smell	
5	Non-toxic	
6	Non-tattooing ink for all skin types	
7	Individually packed	
8	Single Use	
9	Sterile to come with separate 15cm approx. ruler	
10	Suppliers of double ended marker – you will need to apply for both Fine Tip and Regular Lots of you wish to be considered and evaluated for both Lots	
11	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 21 – Tape Measure</u>		Benchtop Evaluation Pass/Fail
1	Tape Measure	
2	Minimum of 100cm	
3	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 22 – Sputum Cup and Lid</u>		Benchtop Evaluation Pass/Fail
1	Sputum Cup and Lid for Cup	
2	Minimum of 100ml	
3	Single Use	
4	Plastic	
5	Packaging must include: Sterile, Single Use, Product Code, Lot No, Expiry date and Conformity Assessment Marking	

<u>Lot 23 – Utility Scissors ‘Tough Cut’</u>		Benchtop Evaluation Pass/Fail
1	Stainless steel blades	
2	Durable plastic handle	
3	Lower blade has a serrated edge and is hooked and rounded at the lip for ease of use	
4	Size must be between 15cm to 20cm	
5	Packaging must include: Product Code, Lot No and Conformity Assessment Marking	

<u>Lot 24 – Emesis Bag</u>		Benchtop Evaluation Pass/Fail
1	Rigid plastic top moulded with no sharp edges	
2	Graduated to show volume of contents	
3	Capacity of 1,000 to 2,000 Millilitres	
4	Means of sealing/security bag to prevent spills	
5	Packaging must include: Product Code, Lot No and Conformity Assessment Marking	